

Il costruttore
The manufacturer
Le fabricant
Der Hersteller

Gefran SPA
Via Sebina, 74
25050 – Provaglio d'Iseo (BS) – Italy

Dichiara sotto la sua esclusiva responsabilità che i prodotti GEFTRAN riportanti la marcatura "UKCA":
Declares on his own responsibility that GEFTRAN products with the "UKCA" mark, series:
Déclare sous sa responsabilité que les produits GEFTRAN avec le marquage "UKCA":
Erklärt hiermit, dass die GEFTRAN - Produkte mit "UKCA" - Kennzeichnung:

Load cells, series

CIR

Con estensione a tutti i modelli descritti dalla seguente sigla di ordinazione:
With extension to all the models described from the following order code:
Avec l'extension à tous les modèles décrits dans le code de commande suivante:
Mit der Erweiterung auf alle Modelle in der folgenden Reihenfolge Code beschrieben:

CIR – A – B – C – D – E

ORDER CODE CONFIGURATION	
Field "A"	B, C, E, M, N
Field "B"	2, 3
Field "C"	M, N
Field "D"	0, Z
Field "E"	XXX if present for customised versions

- sono in accordo con i Regolamenti del Regno Unito
- are manufactured according with United Kingdom Regulations
- ont été fabriqué selon la Réglementation britannique
- mit den britischen Vorschriften übereinstimmen

α) Restriction of the use of certain hazardous substances in Electrical and Electronic Equipment Regulations 2012
β) Electromagnetic Compatibility Regulations 2016


- e che sono state applicate le seguenti (parti/articoli di) norme designate
- and that the following (parts/articles of) designated standards and regulation have been applied
- et que les normes (partes/articles de) désignées suivantes ont été appliqués
- die folgenden bezeichneten Normen Anwendung finden

α) **EN IEC 63000:2018** Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

β) **EN 61326-1:2013** Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

β) **EN 61326-2-3:2013** Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-3: Particular requirements - Test configuration, operational conditions and performance criteria for transducers with integrated or remote signal conditioning

UKCA marking date: 07.06.2022

Date:	BU Manager Paolo Butti 
Gefran SPA Via Sebina, 74 - 25050 Provaglio d'Iseo (BS)	Tel. +39 030 9888 1 - Fax +39 030 9839063 P. IVA 03032420170