

Certificate of Compliance

Certificate Number 260706 - E216851
Report Reference E216851, July 25th, 2006
Issue Date 2006 July 26

Page 1 of 1



Issued to: **Gefran S.p.A.**
Via Sebina 74
I-25050 Provaglio D'Iseo Brescia Italy

This is to certify that representative samples of

Process Control Equipment, Electrical

Models 2400 followed by 0 or 1, followed by 0 or W, followed by 2R, 2D, 4R or 4D, followed by 0,2,3 or P, followed by 0 or 1, may be followed by Y, XX where Y may be 0 or a character between A and Z and XX a number between 00 and 99; Model 2500 followed by 0 or 1, followed by 0 or 1, followed by 0 or 1, followed by 0 or W, followed by 0,2,3 or P follow by 0 or 1, may be followed by Y, XX where Y may be 0 or a character between A and Z and XX a number between 00 and 99.

Have been investigated by Underwriters Laboratories Inc.® in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety: **UL 61010A-1 - Electrical Equipment for Laboratory Use; Part 1: General Requirements**
CSA C22.2 No. 61010.1 - Safety of Electrical Equipment for Measurement, Control, and Laboratory Use

Additional Information: **Electrical Ratings:**
Voltage: 20 - 27 Vac/Vdc or 100 - 240 Vac/Vdc; Frequency: 50/60 Hz;
Power: 20 VA; Digital Inputs : 24 V, max. 5mA; Auxiliary Inputs - Linear DC; 0-20, 4-20 mA, 50 ohms; 10V, Ri>2Mohm; Potentiometer: 1-10 Kohm, 10Vdc

Only those products bearing the UL Listing Mark for the US and Canada should be considered as being covered by UL's Listing and Follow-Up Service meeting the appropriate requirements for US and Canada.

The UL Listing Mark for the US and Canada generally includes: the UL in a circle symbol with "C" and "US" identifiers:  the word "LISTED"; a control number (may be alphanumeric) assigned by UL; and the product category name (product identifier) as indicated in the appropriate UL Directory.

Look for the UL Listing Mark on the product

Issued by: **Kiya Ghamari**
Kiya Ghamari, Associate Project Engineer

UL International Germany GmbH

Any information and documentation provided to you involving UL Mark services are provided on behalf of Underwriters Laboratories Inc.
For questions in Germany, you may call +49 0 6102 3690.

Reviewed by: **Walter Hofmair**
Walter Hofmair, Senior Project Engineer

UL International Germany GmbH