



Certificate of Compliance

Certificate: 70209471

Master Contract: 232831

Project: 70215091

Date Issued: 2019-02-18

Issued to: JUMO GmbH & Co. KG
Moritz-Juchheim-Straße 1
36039 Fulda
GERMANY

Attention: Ulrich Theiss

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US (indicating that products have been manufactured to the requirements of both Canadian and US Standards) or with adjacent indicator 'US' for US only or without either indicator for Canada only



Issued by: Wardell Dumas
Wardell Dumas

PRODUCTS

CLASS – 2252 06 PROCESS CONTROL EQUIPMENT

CLASS – 2252 86 PROCESS CONTROL EQUIPMENT – (Certified to US Standards)

Component type, hydrostatic pressure level probe, model JUMO MAERA S26 (Type 402090), rated 25mA max, 30Vdc max, 10 bar max. Output ratings range between 0-20mA or 0-10V.

Conditions of Acceptability:

1. The device is not to be directly connected to a supply MAINS. The device shall be powered by a separately certified Power Supply Unit with a limited energy electric circuit in accordance with CSA/UL/EN/IEC 61010-1:2010, clause 9.4. Alternatively, a class 2 supply in accordance with CSA223/UL1310 shall be used.
2. Equipment has only been tested for electrical safety. No evaluation of functional safety and performance have been performed.
3. The equipment has been tested for use in ambient temperature of 0°C to 50°C



Certificate: 70209471

Master Contract: 232831

Project: 70209471

Date Issued: 2019-02-18

4. The equipment is intended to be used in wet locations after installation.
5. If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.
6. Equipment has only been tested for use in ordinary locations. No evaluations for use in hazardous locations have been performed.

APPLICABLE REQUIREMENTS

CSA Standards:

CAN/CSA-C22.2 No. 61010-1-12

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Standards:

UL Std. No. 61010-1 (3rd Edition)

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

MARKINGS

See report



Supplement to Certificate of Compliance

Certificate: 70209471

Master Contract: 232831

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
70215091	2019-02-18	Update to report 70209471 for typos and corrections of acceptability.
70209471	2019-01-18	Original cCSAus certification of the JUMO MAERA S26 (Type 402090), a component type, hydrostatic pressure level probe, to C22.2 No. 61010-1-12 / UL 61010-1 (3 rd Edition) under class 2252 06/86.